



Infection Prevention and Control Lapse Report Vital Medical Centre

Initial Report

Premise/facility under investigation

Vital Medical Centre
380 Hunt Club Road
Ottawa, ON
K1V 1C1

Type of premise/facility

Medical Clinic

Date Board of Health became aware of IPAC lapse

2023-12-05

Date of Initial Report posting

2023-12-21

Date of Initial Report update(s)

N/A

How the IPAC lapse was identified

In response to an IPAC complaint, Ottawa Public Health (OPH) inspected the facility and an IPAC lapse was identified.

Summary Description of the IPAC Lapse

Reprocessing of reusable medical equipment:

- Layout of medical device reprocessing area (MDRA)
- Cleaning and disinfection
- Sterilization
- Documentation
- Education of staff

IPAC Lapse Investigation

Did the IPAC lapse involve a member of a regulatory college?

Yes

If yes, was the issue referred to the regulatory college?

The College of Physicians and Surgeons of Ontario has been informed of the investigation.

Were any corrective measures recommended and/or implemented?

On 2023-12-05, the facility was instructed to stop onsite medical device reprocessing using the autoclave and to use single use disposable items.

Please provide further details

The facility is working with OPH to ensure that the MDRA and reprocessing processes are congruent with best practices.

Date any order(s) or directive(s) were issued to the owners/operators (if applicable)

N/A

Initial Report Comments and Contact Information

Any Additional Comments

OPH determined that despite the lapse in IPAC processes, medical instrument cleaning and sterilization was adequate.

If you have any further questions, please contact:

Title: Michelle Wasser, Program Manager
Infection Prevention and Control Inspections and Investigations
Ottawa Public Health

E-mail address: IPAC/PCI@ottawa.ca ,

Phone number: 613-580-2424 ext. 26325

For general updates regarding this investigation, continue to monitor this report.

The results of routine inspections are posted on the [Ottawa Public Health Disclosure website](#).

Final Report

Date of Final Report posting:

2024-08-01

Date of Final Report Update(s)

2024-08-01

Date all corrective measures were confirmed to have been completed

2024-07-18

Brief description of corrective measures taken

On 2024-07-18, OPH confirmed that the clinic has taken the following actions to address the IPAC deficiencies:

- Dedicated reprocessing area
- Procedure for reprocessing that follows IPAC best practices
- Staff completed training in reprocessing

Any Additional Comments

Clinic advised they can resume use of reusable equipment on 2024-07-18.

If you have any further questions, please contact:

Title: Michelle Wasser, Program Manager

Infection Prevention and Control Inspections and Investigations

Ottawa Public Health

E-mail address: IPAC/PCI@ottawa.ca ,

Phone number: 613-580-2424 ext. 26325

For general updates regarding this investigation, continue to monitor this report.

The results of routine inspections are posted on the [Ottawa Public Health Disclosure website](#).